

**“Evaluation of radiation dose reduction strategies in pediatric radiology”**

**Omar Ahmed Alkhairi  
Ahmed Mohammed Abdullah ALAnsari  
HASSAN Jamal Jeddawi  
Mohammed Muslim Salm Alharbi  
Abdalgader Mohammed Faizo  
Rashad Fawzi Rashad Beyari  
Rozan yousef Hamza Mishiddi  
Abdulrahman Mohammed Abdulrahman Aldaoud  
Khaled Abdullah Alturkistani  
Karam adnan wafa  
Ammar Sami Felemban  
Asmaa Abdulrhman ALJIFRI**

### Abstract:

This paper examines the obstacles and methodologies associated with radiation dose reduction in pediatric radiology, emphasizing the impediments to the uniform use of these methodologies. A survey was administered to evaluate the understanding and behaviors of healthcare workers about radiation dose reduction strategies. Essential measures addressed encompass the ALARA principle, age-appropriate imaging modalities, tailored imaging procedures, and the implementation of modern technologies like Automatic Exposure Control (AEC). The survey revealed that although the majority of experts acknowledged the need of reducing radiation doses, obstacles such as reluctance to novel technologies, budgetary limitations, inadequate training, and restricted access to sophisticated equipment were prevalent. The results underscore the necessity for enhanced training, greater accessibility to modern technology, and the cultivation of a culture that emphasizes radiation safety.

**Keywords:** radiation dose reduction, pediatric radiology, ALARA principle, Automatic Exposure Control (AEC), age-appropriate imaging, and healthcare professionals.

### المستخلص:

تدرس هذه الورقة العقبات والمنهجيات المرتبطة بخفض جرعات الإشعاع في الأشعة للأطفال، مع التركيز على العوائق التي تحول دون الاستخدام الموحد لهذه المنهجيات. تم إجراء استطلاع لتقييم فهم وسلوك العاملين في مجال الرعاية الصحية فيما يتعلق باستراتيجيات خفض جرعات الإشعاع. تشمل التدابير الأساسية التي تمت مناقشتها مبدأ ALARA ، وأساليب التصوير المناسبة للعمر ، وإجراءات التصوير المصممة خصيصًا ، وتنفيذ التقنيات الحديثة مثل التحكم التلقائي في التعرض (AEC). كشف الاستطلاع أنه على الرغم من اعتراف غالبية الخبراء بالحاجة إلى خفض جرعات الإشعاع، إلا أن العقبات مثل الإحجام عن التقنيات الجديدة، والقيود المالية، والتدريب غير الكافي، والوصول المحدود إلى المعدات المتطورة كانت سائدة. تؤكد النتائج على ضرورة التدريب المعزز، وزيادة إمكانية الوصول إلى التكنولوجيا الحديثة، وزراعة ثقافة تؤكد على سلامة الإشعاع.

**الكلمات الرئيسية:** خفض جرعات الإشعاع، الأشعة للأطفال، مبدأ ALARA ، التحكم التلقائي في التعرض (AEC) ، التصوير المناسب للعمر، والعاملين في مجال الرعاية الصحية.

**Introduction:**

Pediatric radiology is a specialist field of medical imaging dedicated to the diagnosis and therapy of pediatric medical disorders utilizing imaging technology. It includes several imaging modalities such as X-rays, computed tomography (CT), magnetic resonance imaging (MRI), ultrasound, and nuclear medicine, specifically designed for pediatric patients. Pediatric radiologists possess specialized training to analyze pictures from these technologies, considering the unique anatomical and physiological distinctions between children and adults that might influence both image capture and interpretation (Thukral, 2015).

Children are not simply "small adults" regarding medical imaging. Their bodies are still maturing, and their tissues exhibit heightened sensitivity to radiation, rendering them more susceptible to the possible detrimental effects of ionizing radiation employed in various imaging modalities. Pediatric radiography is essential for identifying a wide array of illnesses in children, including developmental abnormalities and severe trauma. Pediatric radiology prioritizes a meticulous and intentional methodology in diagnostic imaging, reconciling the necessity for precise diagnosis with the obligation to reduce radiation exposure (Portelli, et al.2018).

The cornerstone of pediatric radiology is the formulation and execution of procedures especially aimed at minimizing radiation exposure in children. This is crucial because, in young populations, the dangers associated with radiation exposure—such as possible developmental harm and elevated lifelong cancer risk—are far larger than in adults.

Numerous radiation dose reduction measures have been implemented in pediatric radiology throughout the years to mitigate these problems. This include enhancements in imaging technology, refined scanning techniques, and the adoption of evidence-based practices. Methods include automated exposure control (AEC), reduced-dose CT procedures, and iterative reconstruction techniques have demonstrated efficacy in minimizing radiation exposure while maintaining diagnostic integrity. The efficacy of these techniques in clinical environments is still under continuous assessment (Dudhe, et al.2024).

This study seeks to assess the efficacy of existing radiation dose reduction measures in pediatric radiology via statistical analysis. We will evaluate the effects of various tactics on radiation dose levels and diagnostic accuracy by reviewing data gathered from healthcare facilities utilizing these approaches. Statistical techniques will be utilized to discern patterns, correlations, and prospective enhancements in pediatric imaging procedures. This study's findings will provide critical insights for enhancing radiation safety practices, hence improving health outcomes for pediatric patients.

**Problem Statement:**

The application of ionizing radiation in pediatric radiology is crucial for precise diagnosis and treatment planning; nonetheless, it presents considerable hazards to young children owing to their heightened sensitivity to radiation and the possibility of long-term health repercussions. Pediatric patients are especially sensitive to radiation-induced harm due to their cells' heightened sensitivity to radiation's biological effects and their extended life expectancy, during which adverse consequences may emerge. Notwithstanding the progress in imaging technologies and the implementation of radiation dose reduction strategies, including optimized imaging protocols, low-dose CT scans, and automatic exposure control (AEC), there persists a significant necessity to evaluate the practical efficacy of these strategies in reducing radiation exposure while preserving diagnostic accuracy. Numerous healthcare institutions implement these tactics without a thorough assessment of their efficacy in clinical practice, prompting questions over the real dosage reductions attained and the adequacy of safety measures for young patients.

This study seeks to fill the void in the existing knowledge on radiation dose reduction techniques in pediatric radiology. This study will evaluate current methods, identify barriers to their widespread adoption, and assess their effectiveness in reducing radiation exposure while preserving diagnostic quality, thereby offering essential insights for enhancing patient safety and optimizing radiological practices in pediatric care.

**Study Objectives:**

1. To identify and categorize the radiation dose reduction strategies currently employed in pediatric radiology.
2. To assess the long-term health outcomes associated with reduced radiation exposure in pediatric patients.
3. To determine the barriers to the implementation and consistent use of radiation dose reduction strategies in pediatric radiology.

**Study Significance:**

This study is significant due to its ability to tackle essential issues related to radiation safety in pediatric radiography. Children have heightened sensitivity to ionizing radiation due to their developing tissues, elevated metabolic rates, and extended life expectancy, hence increasing the risk of radiation-induced health disorders, including cancer. As diagnostic imaging remains essential in the medical care of juvenile patients, it is crucial to reduce radiation exposure while preserving diagnostic accuracy. This study will yield significant insights into the efficacy of current radiation dose reduction measures, allowing healthcare practitioners to discern ways that optimally combine patient safety with the necessity for high-quality diagnostic imaging. The project will enhance clinical guidelines and procedures by assessing current practices and identifying hurdles to their adoption, ultimately increasing overall safety standards in pediatric radiology.

The findings will significantly impact policy formulation, education, and training in radiology departments, fostering enhanced knowledge and compliance with optimal techniques for radiation dose reduction. The results may facilitate the establishment of more standardized, evidence-based methodologies for pediatric imaging, minimizing practice variability and providing uniform, safe care across healthcare facilities.

This study may enhance long-term health benefits by reducing radiation exposure hazards in pediatric populations, therefore improving patient outcomes and promoting a culture of safety in pediatric radiology.

### Limitations of the study:

1. **Inconsistent Implementation:** Not all healthcare institutions may have uniformly executed or recorded radiation dose reduction techniques, resulting in incomplete or unequal data.
2. **Access to Proprietary Protocols:** The study may encounter challenges in acquiring access to private imaging procedures or technology utilized by various institutions, hence restricting a thorough assessment of specific practices.
3. **Variability in Professional Practices:** Divergences in the interpretation and implementation of dose reduction methods across radiology practitioners may influence the uniformity and efficacy of these strategies.
4. **Long-term Health Outcome Measurement:** Assessing long-term health outcomes of dose reduction strategies is challenging, as it requires extensive follow-up periods and may be influenced by multiple factors beyond the scope of this study.

### Definition of key terms:

#### 1. Pediatric Radiology:

A medical speciality dedicated to employing imaging modalities, including X-rays, CT scans, and MRI, for the diagnosis and management of health issues in pediatric patients.

#### 2. Ionizing Radiation:

Is a sort of energy emitted by atoms, manifesting as electromagnetic waves (gamma or X-rays) or particles (neutrons, beta, or alpha). The spontaneous decay of atoms is termed radioactivity, and the surplus energy released constitutes a sort of ionizing radiation. Radionuclides are unstable elements that disintegrate and release ionizing radiation, potentially damaging or altering cellular structures and elevating the risk of cancer and other health complications (Mettler, 2012).

#### 3. Radiation Dose Reduction Strategies:

Methods and procedures aimed at decreasing the quantity of ionizing radiation utilized in medical imaging while preserving diagnostic picture quality. These tactics encompass the optimization of imaging protocols, the utilization of modern technology, and the application of the "as low as reasonably achievable" (ALARA) concept.

#### 4. Diagnostic Image Quality:

The precision and clarity of medical images, must be sufficiently good to enable healthcare practitioners to render correct diagnoses while concurrently reducing radiation exposure.

#### 5. Automatic Exposure Control (AEC):

A mechanism in radiological apparatus that autonomously modulates the radiation dosage according to the dimensions and density of the anatomical region being examined, therefore assuring the minimal feasible dose while preserving picture quality (Favazza, et al.2015).

#### 6. Iterative Reconstruction:

A computer technique employed in CT imaging to augment picture quality and decrease radiation exposure by processing raw data via several iterations to diminish noise and boost clarity.

#### 7. As Low as Reasonably Achievable (ALARA):

A radiological safety approach designed to limit radiation exposure by refining imaging techniques and processes to eliminate superfluous doses while maintaining diagnostic quality (Hansson, 2013).

#### 8. Radiology Professionals:

Healthcare practitioners, including radiologists and radiologic technicians, who specialize in employing imaging modalities for diagnostic and therapeutic applications.

### Literature Review:

#### 1. Importance of Radiation Dose Reduction in Pediatrics:

- **Increased Sensitivity to Radiation:** The growing tissues of children exhibit markedly greater sensitivity to ionizing radiation than those of adults. This increased sensitivity results from the accelerated cell division and expansion that transpires during childhood. Rapid cell division heightens the likelihood of radiation disrupting DNA, resulting in mutations and elevating the risk of radiation-induced illnesses. Minimizing radiation exposure in young patients is essential to prevent both acute and long-term health repercussions (Kutanzi, et al.2016).
- **Longer Life Expectancy:** Children possess a longer life expectancy than adults, indicating they have more years for any possible detrimental consequences of radiation exposure to emerge. The prolonged duration heightens the likelihood of incurring long-term problems, including cancer or other radiation-associated health issues. The extended latency period between exposure and the manifestation of radiation-induced disorders underscores the necessity of minimizing radiation doses during imaging procedures conducted in children.
- **Higher Cumulative Exposure Risk:** Pediatric patients frequently necessitate more imaging throughout their lives, whether for continuous treatment, chronic disease monitoring, or emergency evaluations, resulting in a higher cumulative exposure risk. This leads to increased cumulative radiation exposure, potentially exacerbating concerns over time. Strategies for dose reduction are crucial in alleviating these hazards and averting the cumulative effects of many exposures during a child's growth and development (Frush, 2021).

- **Cellular Development:** Children are in a stage of accelerated growth and cellular development, rendering their tissues and organs more susceptible to radiation. The brain, thyroid, and bone marrow are especially susceptible. Any harm during these critical developmental phases may result in enduring consequences for organ function and general health. Minimizing radiation exposure safeguards these vulnerable regions from possible damage (Kutanzi, et al.2016).
- **Necessity for High-Quality Diagnostics:** Despite the inherent hazards of radiation, diagnostic imaging is essential for effective medical care, allowing healthcare professionals to effectively detect, diagnose, and treat illnesses. Preserving high-quality photos is crucial for enabling clinicians to make educated judgments. This underscores the necessity for a meticulous strategy in which dosage reduction is harmonized with the demand for images that offer adequate information for precise diagnosis.

## 2. Radiation Dose Reduction Strategies:

- **Optimization of Imaging Protocols:**

Customizing imaging techniques according to the patient's age, size, and clinical condition is crucial for reducing radiation exposure. Protocol modifications may involve reducing the radiation dosage for particular procedures (such as chest X-rays or CT scans) while preserving diagnostic image quality. Pediatric-specific procedures are formulated to utilize the minimal dose necessary to ensure adequate picture clarity for precise diagnosis (Tsapaki, 2020).

- **Use of Age-Appropriate Imaging Techniques:**

Pediatric patients have particular imaging methods owing to their reduced body size and heightened susceptibility to radiation. Age-appropriate environments guarantee the reduction of radiation exposure while maintaining the integrity of diagnostic images. This entails modifying the exposure parameters according to the child's dimensions, mass, and the specific imaging modality needed (e.g., altering kilovolt (kV) and milliamperage (mA) settings) (Alzen, & Benz-Bohm, 2011).

- **Automatic Exposure Control (AEC):**

AEC systems autonomously modify the radiation dosage according to the patient's physique and the region being examined. This guarantees that the radiation dose is sufficient to generate a diagnostic-quality image, minimizing superfluous exposure. Automatic Exposure Control (AEC) is especially beneficial in pediatric imaging, where variations in body size necessitate dose adjustments to ensure successful imaging while minimizing radiation exposure (Favazza, et al.2015).

- **Iterative Reconstruction Technology:**

Iterative reconstruction is a computer method employed in CT imaging to enhance image quality while minimizing radiation exposure. Iterative reconstruction minimizes picture noise and enables reduced radiation exposure while maintaining diagnostic accuracy through repeated processing of raw data. This method is particularly advantageous in pediatric radiology, where the objective is to reduce dosage while maintaining high-quality images (Hara, et al.2009).

- **Limiting Repeat Examinations:**

A highly effective method to minimize radiation exposure is to refrain from needless repeat imaging. This can be accomplished by guaranteeing that the initial imaging is of superior quality, accurately positioned, and utilizing suitable exposure settings. Establishing quality control protocols, including routine training for imaging personnel, can decrease the necessity for repeat examinations due to substandard image quality or positional inaccuracies.

- **Use of Alternative Imaging Modalities:**

Non-ionizing imaging modalities, such as ultrasound or magnetic resonance imaging (MRI), should be prioritized as alternatives to radiography procedures whenever feasible. These imaging methods are non-radiative and are especially effective for visualizing soft tissues, prevalent in pediatric disorders. MRI can supplant CT scans in numerous situations, particularly when minimizing radiation exposure is a paramount consideration (Bhargava, et al.2013).

- **Shielding:**

Effective shielding methods are crucial to safeguard vulnerable regions of the body from excessive radiation exposure. Lead aprons, thyroid shields, and gonadal shields serve to minimize exposure to essential organs and tissues, especially in young patients. Nonetheless, shielding must be meticulously implemented to prevent obstruction of the area of interest throughout the imaging process (Cheon, et al.2018).

- **Radiologist and Technologist Training:**

The ongoing education and training of radiologists and technicians are essential for enhancing their comprehension of radiation dose reduction principles. Ensuring that imaging experts are knowledgeable and skilled in contemporary dose reduction measures, including improved protocols and sophisticated technologies, can limit overall radiation exposure while preserving diagnostic accuracy.

- **Regular Dose Audits and Monitoring:**

Systematic evaluations of radiation dose levels in imaging processes can pinpoint opportunities for dose reduction while maintaining image quality. Monitoring programs evaluate radiation exposure in pediatric patients, enabling radiology departments to analyze the efficacy of dose reduction techniques and ensure optimal outcomes.



- **Use of Pediatric-Specific Equipment:**

Pediatric-specific imaging apparatus, like diminutive CT scanners and pediatric X-ray machines, can be utilized to optimize radiation dosage. These machines are engineered to emit reduced radiation levels, specifically for smaller, more sensitive individuals, providing a substantial decrease in exposure while maintaining image quality (Borders, et al.2012).

- **Implementation of the ALARA Principle:**

The ALARA (As Low As Reasonably Achievable) approach must be implemented in all pediatric radiology practices. Healthcare practitioners must consistently endeavor to utilize the minimal radiation dose necessary to obtain the required diagnostic image. The ALARA principle promotes ongoing evaluation and enhancement of radiation protection measures (Dudhe, et al.2024).

### **3. Barriers to the Implementation and Consistent Use of Radiation Dose Reduction Strategies in Pediatric Radiology:**

- **Lack of Awareness and Training:**

Numerous healthcare personnel, like as radiologists and technicians, may not possess comprehensive knowledge of the most recent dose reduction measures or may lack sufficient training for their implementation. Inadequate education may lead workers to unintentionally provide elevated radiation doses, particularly if they lack familiarity with sophisticated technology like iterative reconstruction or automated exposure control (AEC) (Thukral, 2015).

- **Inadequate Resources and Equipment:**

Certain healthcare facilities, particularly in resource-constrained environments, may lack access to the most recent pediatric-specific imaging apparatus or sophisticated radiation dose reduction technologies. The absence of specialist equipment, including pediatric CT scanners, sophisticated image processing software, or AEC systems, can impede the successful execution of dose reduction techniques.

- **Financial Constraints:**

The initial expenses associated with upgrading radiography equipment or adopting new radiation dose reduction technologies may be excessive for certain hospitals or clinics. Moreover, educating personnel to utilize these technologies frequently necessitates considerable commitment of time and resources. Financial limitations may hinder or obstruct the implementation of radiation-reducing devices and processes.

- **Resistance to Change:**

Radiologists and imaging technologists may exhibit resistance to altering existing methods, especially when it entails the adoption of new technology or workflows. In certain instances, experts may perceive conventional imaging techniques as adequate, or they may harbor apprehensions over the time necessary to acclimate to new procedures or systems (Christensen, et a.2024).

- **Balancing Image Quality with Dose Reduction:**

A perceived trade-off may exist between minimizing radiation dose and preserving diagnostic image quality. Certain doctors may be concerned that diminishing radiation exposure could jeopardize the clarity or diagnostic efficacy of the images, particularly in intricate pediatric cases where precise diagnosis is essential. This apprehension may hinder the regular use of dose reduction techniques.

- **Lack of Standardized Protocols:**

Standardized, institution-wide strategies for radiation dose reduction in pediatric radiology are frequently absent. Divergence in imaging protocols among various departments or institutions may result in inconsistent use of dose reduction methods. The lack of globally accepted criteria may hinder healthcare providers in determining optimal strategies for individual patient cases.

- **Inconsistent Monitoring and Quality Assurance:**

In the absence of frequent audits and monitoring of radiation dosage levels, evaluating the efficacy of dose reduction techniques might prove challenging. In certain healthcare environments, quality assurance processes may lack the robustness necessary to monitor the continuous application of radiation dose reduction technology. In the absence of regular feedback and assessment, it is difficult to guarantee the constant implementation of dosage reduction techniques.

- **Pressure to Achieve Quick Diagnostic Results:**

In urgent or rapid clinical environments, there may be pressure to get swift results, perhaps resulting in the prioritizing of speed over the optimization of radiation dosage. Technologists or physicians may be disinclined to adhere to established protocols in the face of urgent demands for fast diagnostic results, resulting in increased radiation exposure.

- **Clinical and Institutional Culture:**

Some healthcare facilities may not emphasize radiation safety or dose reduction in their culture. If dose reduction methods are not regarded as essential to patient care, or if leadership fails to prioritize radiation safety, staff may not consistently comply with these practices. A culture that neglects patient safety and radiation protection can hinder the overall efficacy of dose reduction initiatives (Whitebird, et al.2021).

#### ○ **Technological Limitations and Interoperability Issues:**

Certain radiology systems may lack full compatibility with contemporary dose reduction technology, resulting in challenges when incorporating these systems into established workflows. Furthermore, antiquated systems may lack compatibility with newer functionalities like automatic dose modulation or picture reconstruction techniques, hence complicating the attainment of uniform radiation dose reductions in pediatric patients (Iroju, et al.2013).

#### **4. Progress in Imaging Technology:**

Sophisticated Imaging Apparatus Recent developments in imaging technology have considerably facilitated the lowering of radiation dosage.

- **Digital Radiography:** Digital radiography (DR) systems employ sophisticated detection technology that necessitates reduced radiation exposure relative to conventional film-based systems. Digital Radiography (DR) devices provide superior image quality at reduced radiation doses due to enhanced sensitivity and dynamic range (Bansal, 2006).
- **Low-dosage CT Scanners:** Contemporary CT scanners with iterative reconstruction methods offer substantial dosage reduction while preserving diagnostic image quality. Methods like adaptive statistical iterative reconstruction (ASIR) and model-based iterative reconstruction (MBIR) have demonstrated the capability to decrease radiation exposure by as much as 60% relative to traditional techniques (McLeavy, et al.2021).
- **Advancements in fluoroscopy:** such as pulsed fluoroscopy and automated dose control systems, contribute to the reduction of radiation exposure. Pulsed fluoroscopy minimizes continuous exposure duration, and automatic dose control modulates the radiation dose according to clinical requirements.

#### **Previous Studies:**

According to (Zacharias, et al.2013). The advent of MDCT has augmented the application of CT in pediatric radiology, accompanied by apprehensions regarding radiation sequelae. This article examines fundamental principles for reducing radiation dose, the essential physics influencing radiation exposure, and particular CT-integrated dose-reduction methods aimed at the pediatric demographic. This paper aims to deliver a thorough examination of contemporary research about CT dose reduction techniques, their constraints, and prospective advancements, particularly in relation to the pediatric demographic. The discourse will commence with overarching factors contributing to radiation dose reduction, then addressing certain technical attributes that affect the radiation dose.

The study by (Ngo, et al. 2018) examines realistic methods to reduce radiation exposure during pediatric CT scans. The authors present an extensive array of solutions designed to minimize dosage while preserving diagnostic image quality. The research underscores the necessity of customizing CT procedures to the distinct attributes of pediatric patients, encompassing size and age, and accentuates the adoption of sophisticated imaging technologies like automated exposure control (AEC) and iterative reconstruction. Detailed discussions encompass techniques such as improving scanning parameters, using dose-modulation technologies, and circumventing superfluous repeat scans. The authors also discuss the difficulties encountered by institutions in implementing and standardizing these dose-reduction measures. The study functions as a pragmatic resource for radiologists and healthcare institutions seeking to improve patient safety via efficient radiation management protocols in pediatric imaging. According to (Costello et al. 2013) aim to investigate several facets of ionizing radiation exposure in individuals subjected to CT exams via clinical scenarios. The review examines imaging suitability, optimal radiation dosages, related cancer risks, and essential dose reduction techniques. CT scans, although crucial for their diagnostic use, contribute substantially to radiation exposure associated with medical imaging, underscoring the necessity for a balance between advantages and hazards. The authors emphasize that medical professionals must comprehend the advantages and potential risks of medical radiation and be proficient in strategies to reduce radiation doses, including optimizing imaging protocols, utilizing automatic exposure control (AEC), and implementing iterative reconstruction techniques. Instructing healthcare personnel on these tactics is essential for safeguarding patient safety and fostering appropriate imaging practices that minimize needless exposure while maintaining diagnostic precision.

#### **Methodology:**

##### **1. Study Design:**

The research used a descriptive methodology to examine the efficacy of radiation dose reduction measures in pediatric radiology. This methodology is appropriate for examining the correlation between dose reduction strategies and the quality of diagnostic results. A descriptive study approach facilitates a comprehensive examination of existing solutions, the obstacles encountered in their consistent use, and their effect on reducing radiation exposure in children. This corresponds with the idea articulated by (Kemp, et al.2018), wherein descriptive analytical approaches offer a systematic approach for precisely delineating the subject and delivering interpretable results.

##### **2. Research Method:**

The research technique denotes a systematic framework for data collection, analysis, and interpretation, guaranteeing that the study is structured, dependable, and valid (Walliman, 2021). The research utilizes a quantitative technique to assess the effectiveness of several dose reduction measures in pediatric radiology. This quantitative methodology facilitates the acquisition of measurable data about the effects of various tactics, including automated exposure control (AEC) and iterative reconstruction, hence enabling statistical analysis of outcomes associated with dose reduction.

##### **3. Study Population:**

The research population consists of healthcare professionals engaged in pediatric radiography, including radiologists,

radiologic technicians, and medical physicists, who execute and supervise dose reduction protocols. The demographic is selected for their direct involvement with imaging techniques and radiation safety precautions in pediatric patients. This

emphasis is crucial to guarantee that the results authentically represent actual practices and the obstacles encountered in implementing these techniques. The sample size will comprise 100 medical professionals employed in pediatric radiology departments at various hospitals and imaging facilities.

#### 4. Data collection:

This research entails the systematic collecting of quantitative data concerning the application and results of several radiation dose reduction technologies. This may encompass surveys, structured questionnaires, and data from imaging records that chronicle dosage levels before to and subsequent to the application of certain reduction approaches. The objective is to get dependable, accurate data to assess the efficacy of dose reduction measures and pinpoint any deficiencies or obstacles in their implementation. The precise collection of data is essential for the study's validity, guaranteeing that the analysis yields significant results on optimal techniques for minimizing radiation exposure in pediatric radiology.

##### 4.1 Secondary Sources:

Secondary data aids researchers in developing research questions, improving understanding of the study topic, and broadening knowledge of the subject area. Additionally, it assists in recognizing relevant research approaches and establishes a robust foundation for the succeeding stages of the study. Secondary data not only contextualizes primary data but also enhances swift understanding and interpretation (Christensen, et al.2011). Therefore, we conduct a comprehensive evaluation of all relevant research articles.

Secondary sources, such as books and articles, have been employed for data collection.

##### 4.2 Primary Sources:

Primary sources are vital elements of data collection that provide unique, firsthand information or evidence directly relevant to the research topic. These sources are essential as they offer raw data that remains unexamined, unabridged, and uninfluenced by previous researchers or middlemen.

To achieve the research objectives, data were collected by the distribution of a questionnaire to the study population.

#### 5. Data Analysis:

The term "data analysis" denotes the systematic process of reviewing, cleaning, changing, and interpreting obtained data to draw conclusions, address research questions, or assess hypotheses. Currently, researchers utilize several statistical and computational methods to examine the collected data.

This research employed SPSS for the statistical analysis of the questionnaire data.

#### Results:


##### 1. Demographic Questions:

##### 1.1 Gender:

The following table on the gender distribution of the study sample reveals that males represent 70%, and females comprise 30%.

*Table 1: Gender*

		Frequency	Percent	Valid Percent	Cumulative Percent
Valid	Male	70	70.0	70.0	70.0
	Female	30	30.0	30.0	100.0
	Total	100	100.0	100.0	






## 1.2 Age:

It is evident from the following table regarding the distribution of the study sample according to age, that the highest percentage is (35 - 44 years) with 30%, followed by (25–34 years) with a percentage of 16%, (15–24 years) with a percentage of 25% (45 - 54 years) with a percentage of 21% and (More than 55 years) with a percentage of 8%.

*Table 2:Age*

Age		Frequency	Percent	Valid Percent
Valid	15–24 years	25	25	25
	25–34 years	16	16	16
	35 - 44 years	30	30	30
	45 - 54 years	21	21	21
	More than 55 years	8	8	8
Total		100	100.0	100.0



## 1.3 Professional Role Distribution:

The table showing the distribution of the study sample by professional role reveals that Radiologists represent 35% of the total, followed by Radiologic Technicians at 45%, and Medical Physicists at 20%. The cumulative percentage ensures full representation, totaling 100%.

*Table 3*

Professional Role					
		Frequency	Percent	Valid Percent	Cumulative Percent
Valid	Radiologists	35	35	35	35
	Radiologic Technicians	45	45	45	80
	Medical Physicists	20	20	20	100
	Total	100	100.0	100.0	

### 1. Radiation Dose Reduction Strategies:

- ✓ Statement “The ALARA principle ensures the minimal radiation dose is used, promoting ongoing radiation protection improvements.” came in the first place with an arithmetic mean of 4.22 and a standard deviation of .675. Therefore, the direction of the responses of the study sample is Agree.
- ✓ Statement “Age-appropriate imaging methods adjust settings like kV and mA to reduce radiation for pediatric patients. “came in the second order, with a mean of 4.21 and a standard deviation of .832. Therefore, the direction of the responses of the study sample is Agree.

- ✓ Statement “Customizing imaging techniques based on a pediatric patient's age, size, and condition helps reduce radiation exposure while maintaining image quality.” mean of 4.15 and a standard deviation of .687. Therefore, the direction of the responses of the study sample is Agree.
- ✓ Statement “Automatic Exposure Control (AEC) adjusts radiation dose based on the patient's size and the area being examined, minimizing unnecessary exposure “in the fourth rank came with an arithmetic mean of 3.89 and a standard deviation of .751. Therefore, the direction of the responses of the study sample is neutral.

It was apparent from the table that the most effective radiation dose reduction strategies in pediatric radiology, with the highest mean score, was the ALARA principle, ensuring minimal radiation use with ongoing protection improvements. Following closely were age-appropriate imaging methods and customizing techniques based on a patient's size and condition, both showing strong agreement among the study sample. However, the statement regarding Automatic Exposure Control (AEC) was ranked fourth, with a neutral response indicating less consensus.

**Table 4 Descriptive Statistics of Radiation Dose Reduction Strategies**

Descriptive Statistics				
	Mean	Std. Deviation	p-value	arrangement
Customizing imaging techniques based on a pediatric patient's age, size, and condition helps reduce radiation exposure while maintaining image quality.	4.15	.687	0.001	3
Age-appropriate imaging methods adjust settings like kV and mA to reduce radiation for pediatric patients.	4.21	.832	0.320	2
Automatic Exposure Control (AEC) adjusts radiation dose based on the patient's size and the area being examined, minimizing unnecessary exposure.	3.89	.751	0.121	4
The ALARA principle ensures the minimal radiation dose is used, promoting ongoing radiation protection improvements.	4.22	.675	0.603	1
Radiation Dose Reduction Strategies	4.0680	.36979		

## 2. Barriers to the Implementation and Consistent Use of Radiation Dose Reduction Strategies:

- ✓ Statement “Resistance to adopting new technologies or workflows may arise among radiologists and technologists, particularly when conventional methods are perceived as sufficient.” came in the first place with an arithmetic mean of 4.24 and a standard deviation of .712. Therefore, the direction of the responses of the study sample is Agree.
- ✓ Statement “Financial constraints often prevent hospitals from upgrading their radiography equipment or investing in new radiation dose reduction technologies.” came in the second place with an arithmetic mean of 4.12 and a standard deviation of .844. Therefore, the direction of the responses of the study sample is Agree.
- ✓ Statement “Many healthcare personnel may lack adequate training on the latest dose reduction techniques, leading to unintentional higher radiation doses. “came in the third order, with a mean of 4.09 and a standard deviation of .911. Therefore, the direction of the responses of the study sample is Agree.
- ✓ Statement “Limited access to advanced pediatric imaging equipment and dose reduction technologies can hinder effective radiation management in healthcare facilities.” came in the fourth order, with an arithmetic mean of 3.95 and a standard deviation of .880. Therefore, the direction of the responses of the study sample is Agree.

It was clear from the table that the perceived barriers to implementing radiation dose reduction strategies include resistance to adopting new technologies, financial constraints, and inadequate training on the latest techniques. It was also evident that limited access to advanced pediatric imaging equipment and dose reduction technologies can hinder effective radiation management in healthcare facilities.

**Table 5 Descriptive Statistics of Barriers to the Implementation and Consistent Use of Radiation Dose Reduction Strategies**

Descriptive Statistics				
	Mean	Std. Deviation	p-value	arrangement
Many healthcare personnel may lack adequate training on the latest dose reduction techniques, leading to unintentional higher radiation doses.	4.09	.911	0.000	3
Limited access to advanced pediatric imaging equipment and dose reduction technologies can hinder effective radiation management in healthcare facilities.	3.95	.880	0.292	4
Financial constraints often prevent hospitals from upgrading their radiography equipment or investing in new radiation dose reduction technologies.	4.12	.844	0.009	2
Resistance to adopting new technologies or workflows may arise among radiologists and technologists, particularly when conventional methods are perceived as sufficient.	4.24	.712	0.000	1
Barriers to the Implementation and Consistent Use of Radiation Dose Reduction Strategies	4.0575	.51720		

### Recommendations:

- **Augmented Training Initiatives:** To address the issue of inadequate training, it is advisable for healthcare institutions to provide ongoing educational programs for radiologists and technicians. These initiatives should emphasize contemporary dose reduction methodologies, encompassing iterative reconstruction technology and Automatic Exposure Control (AEC) systems.
- **Investment in Sophisticated Equipment:** Healthcare facilities must prioritize investments in modern imaging equipment specifically for pediatrics and technology for reducing radiation doses. Obtaining funds or partnering with governmental entities and non-profit groups might alleviate the financial limitations that obstruct access to these services.
- **Promoting a Culture of Radiation Safety:** It is important to foster a working culture that emphasizes radiation safety. Radiation safety may be integrated into institutional policies and leadership initiatives. Establishing a conducive atmosphere for the integration of new technology will alleviate opposition to change.
- **Routine Audits and Oversight:** Establishing routine audits and monitoring systems will provide continuous assessment of radiation dosage levels. These systems must incorporate feedback loops to evaluate the efficacy of dose reduction strategies and pinpoint areas for enhancement.

### Conclusion:

The study verifies that, despite considerable progress in formulating and executing radiation dose reduction measures in pediatric radiology, substantial obstacles to their uniform use persist. Opposition to new technology, budgetary constraints, insufficient training, and limited access to modern equipment persist in hindering the extensive adoption of these tactics. Nevertheless, via focused interventions including improved training, investment in equipment, standardized processes, and the promotion of a radiation safety culture, these obstacles can be alleviated. By tackling these problems, healthcare institutions may enhance the protection of pediatric patients while preserving superior diagnostic image quality.

## References:

- Alzen, G., & Benz-Bohm, G. (2011). Radiation protection in pediatric radiology. *Deutsches Ärzteblatt International*, 108(24), 407.
- Bansal, G. J. (2006). Digital radiography. A comparison with modern conventional imaging. *Postgraduate medical journal*, 82(969), 425-428.
- Bhargava, R., Hahn, G., Hirsch, W., Kim, M. J., Mentzel, H. J., Olsen, Ø. E., ... & Vazquez, E. (2013). Contrast-enhanced magnetic resonance imaging in pediatric patients: review and recommendations for current practice. *Magnetic resonance insights*, 6, MRI-S12561.
- Borders, H. L., Barnes, C. L., Parks, D. C., Jacobsen, J. R., Zhou, Y., Hasselquist, B. E., & Betz, B. W. (2012). Use of a dedicated pediatric CT imaging service associated with decreased patient radiation dose. *Journal of the American College of Radiology*, 9(5), 340-343.
- Cheon, B. K., Kim, C. L., Kim, K. R., Kang, M. H., Lim, J. A., Woo, N. S., ... & Kim, J. H. (2018). Radiation safety: a focus on lead aprons and thyroid shields in interventional pain management. *The Korean journal of pain*, 31(4), 244-252.
- Christensen, B. M., Bjällmark, A., Ndipen, I. M., Afram, S. S., & Bazzi, M. (2024). Barriers to radiographers' use of radiation safety principles: A qualitative perspective. *Journal of Medical Radiation Sciences*.
- Christensen, L. B., Johnson, B., Turner, L. A., & Christensen, L. B. (2011). Research methods, design, and analysis.
- Costello, J. E., Cecava, N. D., Tucker, J. E., & Bau, J. L. (2013). CT radiation dose: current controversies and dose reduction strategies. *American journal of roentgenology*, 201(6), 1283-1290.
- Dudhe, S. S., Mishra, G., Parihar, P., Nimodia, D., & Kumari, A. (2024). Radiation dose optimization in radiology: a comprehensive review of safeguarding patients and preserving image fidelity. *Cureus*, 16(5).
- Favazza, C. P., Yu, L., Leng, S., Kofler, J. M., & McCollough, C. H. (2015). Automatic exposure control systems designed to maintain constant image noise: effects on computed tomography dose and noise relative to clinically accepted technique charts. *Journal of computer assisted tomography*, 39(3), 437-442.
- Frush, D. (2021). The cumulative radiation dose paradigm in pediatric imaging. *The British Journal of Radiology*, 94(1126), 20210478.
- Hansson, S. O. (2013). ALARA: What is reasonably achievable?. In *Radioactivity in the Environment* (Vol. 19, pp. 143-155). Elsevier.
- Hara, A. K., Paden, R. G., Silva, A. C., Kujak, J. L., Lawder, H. J., & Pavlicek, W. (2009). Iterative reconstruction technique for reducing body radiation dose at CT: feasibility study. *American Journal of Roentgenology*, 193(3), 764-771.
- Iroju, O., Soriyan, A., Gambo, I., & Olaleke, J. (2013). Interoperability in healthcare: benefits, challenges and resolutions. *International Journal of Innovation and Applied Studies*, 3(1), 262-270.
- Kemp, S. E., Ng, M., Hollowood, T., & Hort, J. (2018). Introduction to descriptive analysis. *Descriptive analysis in sensory evaluation*, 1-39.
- Kutanzi, K. R., Lumen, A., Koturbash, I., & Miousse, I. R. (2016). Pediatric exposures to ionizing radiation: carcinogenic considerations. *International journal of environmental research and public health*, 13(11), 1057.
- McLeavy, C. M., Chunara, M. H., Gravell, R. J., Rauf, A., Cushnie, A., Talbot, C. S., & Hawkins, R. M. (2021). The future of CT: deep learning reconstruction. *Clinical radiology*, 76(6), 407-415.
- Mettler, F. A. (2012). Medical effects and risks of exposure to ionising radiation. *Journal of Radiological Protection*, 32(1), N9.
- Ngo, A. V., Winant, A. J., Lee, E. Y., & Phillips, G. S. (2018, April). Strategies for reducing radiation dose in CT for pediatric patients: how we do it. In *Seminars in Roentgenology* (Vol. 53, No. 2, pp. 124-131). WB Saunders.
- Portelli, J. L., McNulty, J. P., Bezzina, P., & Rainford, L. (2018). Benefit-risk communication in paediatric imaging: what do referring physicians, radiographers and radiologists think, say and do?. *Radiography*, 24(1), 33-40.
- Thukral, B. B. (2015). Problems and preferences in pediatric imaging. *Indian Journal of Radiology and Imaging*, 25(04), 359-364.
- Tsapaki, V. (2020). Radiation dose optimization in diagnostic and interventional radiology: Current issues and future perspectives. *Physica Medica*, 79, 16-21.
- Walliman, N. (2021). *Research methods: The basics*. Routledge.
- Whitebird, R. R., Solberg, L. I., Bergdall, A. R., López-Solano, N., & Smith-Bindman, R. (2021). Barriers to CT dose optimization: the challenge of organizational change. *Academic radiology*, 28(3), 387-392.
- Zacharias, C., Alessio, A. M., Otto, R. K., Iyer, R. S., Philips, G. S., Swanson, J. O., & Thapa, M. M. (2013). Pediatric CT: strategies to lower radiation dose. *American Journal of Roentgenology*, 200(5), 950-956.

## Appendix

### Demographic Data Section

#### Age:

- 15-24
- 25-34
- 35-44
- 45-54
- More than 55 years

#### Gender:

- Male
- Female

#### Professional Role:

- Radiologists
- Radiologic Technicians
- Medical Physicists

1 = Strongly Disagree	2 = Disagree	3 = Neutral	4 = Agree	5 = Strongly Agree
-----------------------	--------------	-------------	-----------	--------------------

<b>Radiation Dose Reduction Strategies</b>						
NO.	Items	1	2	3	4	5
1	Customizing imaging techniques based on a pediatric patient's age, size, and condition helps reduce radiation exposure while maintaining image quality.					
2	Age-appropriate imaging methods adjust settings like kV and mA to reduce radiation for pediatric patients.					
3	Automatic Exposure Control (AEC) adjusts radiation dose based on the patient's size and the area being examined, minimizing unnecessary exposure.					
4	The ALARA principle ensures the minimal radiation dose is used, promoting ongoing radiation protection improvements.					

<b>Barriers to the Implementation and Consistent Use of Radiation Dose Reduction Strategies</b>						
NO.	Items	1	2	3	4	5
1	Many healthcare personnel may lack adequate training on the latest dose reduction techniques, leading to unintentional higher radiation doses.					
2	Limited access to advanced pediatric imaging equipment and dose reduction technologies can hinder effective radiation management in healthcare facilities.					
3	Financial constraints often prevent hospitals from upgrading their radiography equipment or investing in new radiation dose reduction technologies.					
4	Resistance to adopting new technologies or workflows may arise among radiologists and technologists, particularly when conventional methods are perceived as sufficient.					